# UVC-19: STERI-WAND



**CPS-UVC-19-SW-PPE** 

## **Operating Instructions**

 Open battery cover and load 4 AA batteries.
Remove the lampshade and press POWER button for 3 seconds to turn it on. UV light and timer can be turned off if the POWER button is pressed again and held for 3 seconds.
Timer setting is activated after pressing the timer button over 3 seconds, 5-minute, 15-minute, 30-minute and continuous working modes can be selected with timing button.

4. 10mm-40mm (1cm to 4cm) distance is suggestedbetween disinfection target surface and the product.5. 2-3 scans are needed for extra dirty areas.

6. Users should also ensure these sterilization areas are fully covered by UV rays. Better disinfection can be achieved in shorter distance.

7. Power it off and cover the lampshade well to protect bulbs from damage after sterilization.

\*\* Please check out the videos of this product on our website and youtube channel for more guidance on use.

#### **Safety Features**

LED LIGHTING

1. Special PCD lampshade can protect users from exposing to UVC-rays.

2. This UVC Sterilizing Wand will shut down automatically when it is turned over.

3. The portable sterilization wand will stop working automatically after setting time.

4. 3 second delay start systems.

#### Waste Disposal

1. Used batteries should be properly processed or recycled according to local laws in order to protect the environment.

2. Mercury inside the UV bulb of this product. Bulb should be put into a plastic bag and properly processed.

3. Used bulb or bulb fragments are suggested to be sent to waste disposal center.

#### CAUTION

1. UVC Sterilizing Wand is not waterproof, please keep it away from water.

2. Keep out of reach of children.

3. UV rays can hurt your eyes or skin, a proper distance should be maintained.

4. This wand is for surface disinfection.

5. Will not work if a bulb is broken, battery is dead or installed wrong, and button failures.

6. Cannot be cleaned with detergent, gasoline, glass cleaner, or any other house hold cleaners.

7. If batteries are low, multiple scans are suggested for performance until changed out.

The information and recommendations contained in this publication are based upon data collected by CPS LED Lighting and are believed to be correct. However, no guarantee or warranty of any kind, expressed or implied, is made with respect to the information contained herein. Specifications and information are subject to change without notice. Although every precaution has been taken in the preparation of this material, CPS LED Lighting assumes no responsibility for errors or omissions. Neither is any liability assumed for damages resulting from the use of the information contain herein.

### CPS LED LIGHTING